



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION
INSPECTORATE EAST AFRICA LIMITED

ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/44**

PERMANENT ADDRESS OF LABORATORY

P.O. Box 42327-80100

TreTra Suites Building, Plot No. 409,

Section No. XXI/MI Off Mnazi Mmoja Road,

Mombasa, Kenya.

Tel: +254 (0)714 323930; (0)723 282499; (041)2317318; (041)2228259

Email Address: admin@inspectorate-ea.com

Laboratory Contact Person(s)

Kyalo Ilunga

Approved by: _____

KENAS CEO/Authorized Representative

Date: 18-June-2020



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION - TESTING

No.	Testing Field	Type of Test	Test Method	Test Object / Matrix	Test Parameter	Measurement Range	Test Location
1.	Chemistry	Volatility	ASTM D 86	Automotive Gas Oil	Distillation	Maximum 36 - 400°C	INSPECTORATE EAST AFRICA LIMITED TESTING LABORATORY
		Gravimetric	ASTM D 1298	Automotive Gas Oil	Density @ 20°C	0.6000 – 1.2500 Kg/L	
		Volatility	ASTM D 93	Automotive Gas Oil	Flash Point, (Pensky Martens Closed Cup °C)	19°C - 400°C	

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Assessment	14-SEP-2018	14-SEP-2018	13-SEP-2022

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past.

Approved by: *Jembo*
KENAS CEO/Authorized Representative

Date: 18-June-2020