



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenyaaccreditaton.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

ESTEC LIMITED

ISO/IEC 17025:2017

Calibration Laboratory Number: **KENAS/CL/008**

PERMANENT ADDRESS OF LABORATORY

ESTEC LIMITED
BUNGOMA ROAD, OFF BARICHO ROAD
P. O. BOX 12143-00400
NAIROBI, KENYA
Tel: 0206537709 / 10 / 0206550876
Email: info@estec.com

Technical Manager:

Mr. Tom Ouma

Technical Signatories:

Mr. Geoffrey Koech

Date of expiry: 22-Feb-2024

Approved by: _____

Date: 21-Jan-2021

KENAS CEO/Authorized Representative



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenyaaccreditaton.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

No	Calibrati on Field	Calibration Object	Measur- and	Laboratory Calibration Method	Standards	Range	Calibration and Measurement Capability (CMC) (Measurement Point / Range ± expanded uncertainty) K = 2	Calibration Site
1	Mass	Non automatic weighing instruments	mass	In-house: ESTEC/BCP/01 Reference Methods: a) EA-4/02, b), EURAMET cg 18:2015	Standard masses Class F1 Traceability BS/MET/2/3/ 88/765 Class M2 Traceability BS/MET/2/3/ 88/764	1mg – 10 kg 5kg - 500 kg	± 0.0008 g - 0.077 g ± 0.008 kg - 0.2 kg	Field calibration

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	22-Feb-2012	22-Feb-2012	21-Feb-2015
1 st Re-Accreditation	22-Feb-2015	22-Feb-2015	21-Feb-2018
2 nd Re-Accreditation	22-Feb-2018	22-Feb-2018	21-Feb-2021
3 rd Re-Accreditation	22-Feb-2021	22-Feb-2021	21-Feb-2025

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past.

Approved by:  _____
KENAS CEO/Authorized Representative

Date: 21-Jan-2021