



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION
HOPE LABS GROUP LIMITED

ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/72**

PERMANENT ADDRESS OF LABORATORY

Ex Dugsi Sare – Hodan District

Trabuna Road, Mogadishu – Somalia

Tel: Hotline: 652541; Mobile: +252 618693774

Email Address: info@Hopelabs.so

Website: www.hopelabs.so

Laboratory Contact Person

Dr. Abdiladif, Hussein

Approved by:


KENAS CEO/Authorized Representative

Date: 03-February-2022



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION – TESTING

No.	Testing Field	Type of Test	Test Method	Test Object / Matrix	Test Parameter	Measurement Range	Test Location		
1.	Microbiology	Enumeration	ISO 4833-1	Food, swabs & Feeds	Total plate count	≥10 cfu/g	Microbiology Laboratory		
			ISO 4832		Total coliform count	≥10 cfu/g			
			ISO 16649-2		<i>Escherichia coli</i>	≥10 cfu/g			
			ISO 6888-2		<i>Staphylococcus aureus</i>	≥10 cfu/g			
			ISO 11290-2		<i>Listeria spp & monocytogenes</i>	≥10 cfu/g			
			ISO 11290-1		<i>Listeria spp & monocytogenes</i>	Qualitative			
			ISO 21527-1		Yeast & Moulds	≥10 cfu/g			
			ISO 21527-2		Yeast & Moulds	≥10 cfu/g			
			Detection		ISO 6579-1	<i>Salmonella spp</i>		Qualitative	
					ISO 21872-1	<i>Vibrio spp</i>		Qualitative	
					ISO 21567	<i>Shigella spp.</i>		Qualitative	
			Enumeration		ISO 21528-2	Water		<i>Enterobactereacea</i>	≥10 cfu/g
					ISO 6222			Total Plate Count	>1cfu/ml
					ISO 9308-1			Total coliform count	<100 cfu/100ml
		ISO 9308-1		<i>E. coli</i>	<100 cfu/100ml				
		ISO 7899-2		<i>S. faecalis</i>	<100 cfu/100ml				
		ISO 16266		<i>P. aeruginosa</i>	<100 cfu/100ml				
		ISO 6461-2		<i>Sulphite reducing clostridia</i>	<100 cfu/100ml				
		ISO 14189		<i>Clostridium perfringens</i>	<100 cfu/100ml				
		Detection	ISO 19250	Water	<i>Salmonela spp</i>	Qualitative			
			USP 62	Pharmaceuticals and related products	<i>Bile Tolerant Gram-Negative bacteria</i>	Qualitative			
		Enumeration	<i>E. coli</i>						
			<i>Salmonella</i>						
<i>P. aeruginosa</i>									
<i>S. aureus</i>									
<i>Clostridia</i>									
USP 61	<i>C. albicans</i>								
USP 61	Total aerobic microbial count	>10cfu/g							
USP 61	Total yeasts & molds count	>10cfu/g							
2.	Chemistry	Titration	USP, BP, Eur, Ph. Int. monographs	Pharmaceuticals and related products	Assay	90% to 120%	Chemistry Laboratory		
		Physical			Conductivity	0.1 µS/cm			
					Loss on Drying	0 to 100%			
					Disintegration	As per specific monographs			
					pH	1 to 12			
		Polarimetry	Identification	Qualitative					
		Photometric	HLG/LIM/002	Food, Feeds & grains	Total Aflatoxins	0.1ppb			
		Gravimetric	ISO 6496		Moisture content	>1%w/w			
			ISO 5983-1		Crude protein	>0.5%w/w			
		AOAC 927.05	AOAC 930.30	Food	Moisture content	>0.01% m/m			
					Ash content	>0.1% m/m			
		Kjehdahl	ICARDA 5.9.1	Envi &Agric	Total Nitrogen	>1			
		Potentiometric	ICARDA 5.2		Potable & waste water	pH	1-14		
			APHA 4500H+ B	1-14					



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Assessment	03-FEB-2022	03-FEB-2022	02-FEB-2026

This schedule is issued subject to the terms and conditions of KENAS Accreditation.

Approved by:



KENAS CEO/Authorized Representative

Date: 03-February 2022

