



**KENYA ACCREDITATION SERVICE**

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: [info@kenas.go.ke](mailto:info@kenas.go.ke) Web: [www.kenas.go.ke](http://www.kenas.go.ke)

**SCHEDULE OF ACCREDITATION**

**INSPECTORATE EAST AFRICA LIMITED**

**ISO/IEC 17025:2017**

Testing Laboratory Number: **KENAS/TL/044**

**PERMANENT ADDRESS OF LABORATORY**

P.O Box 42327-80100

TreTra Suites Building, Plot No. 409,  
Section No. XXI/MI Off Mnazi Mmoja Road,  
Mombasa, Kenya

**Tel: +254 (0)714323930; (0)723282499; (041)2317318;(041)2228259**

**Email Address: [admin@inspectorate-ea.com](mailto:admin@inspectorate-ea.com)**

**Website: [www.inspectorate-ea.com](http://www.inspectorate-ea.com)**

---

**Laboratory Contact Person**

Kyalo Ilunga

Approved by: \_\_\_\_\_

*[Signature]*  
**KENAS CEO/Authorized Representative**

**Date: 28-October-2021**

*[Signature]*  
**2021-10-02**



### KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: [info@kenas.go.ke](mailto:info@kenas.go.ke) Web: [www.kenas.go.ke](http://www.kenas.go.ke)

#### SCHEDULE OF ACCREDITATION

S/No.	Testing Field	Type of Test	Test Method	Test Object / Matrix	Test Parameter	Measurement Range	Test Location
1.	Chemistry	Volatility	ASTM D86	Automotive Gas Oil and Jet A-1	Distillation	36°C to 400°C	INSPECTORATE EAST AFRICA LIMITED TESTING LABORATORY
		Volatility	ASTM D93	Automotive Gas Oil	Flash Point, (Pensky Martens Closed Cup °C)	19°C to 400°C	
		Gravimetric	ASTM D1298	Automotive Gas Oil	Density @20°C	0.6000 to 1.2500 Kg/L	
				Jet A-1	Density @15°C	775 to 840 Kg/m <sup>3</sup>	

#### Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Assessment	14-SEP-2018	14-SEP-2018	13-SEP-2022

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past.

Approved by: \_\_\_\_\_

*[Signature]*  
KENAS CEO/Authorized Representative

Date: 28-October 2022

*[Signature]*  
2021-10-02