



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

ABSOLUTE SCALES LIMITED

ISO/IEC 17025:2017

Calibration Laboratory Number: **KENAS/CL/42**

PERMANENT ADDRESS OF LABORATORY

Kings Godown No. 5, Road C, Enterprise Road,
Nairobi, Kenya

Tel: +254 20 3599661/3
Mob: +254 710 600 402 / 733 697 066

Email Address: info@absolutescales.co.ke

Laboratory Contact Person

Ritesh Patel

Approved by: _____

KENAS CEO/Authorized Representative

Date: 07-Jan-2022



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION - CALIBRATION

No	Calibration Field	Calibration Object	Measurand	Laboratory Calibration Method	Standards	Range	Calibration and Measurement Capability (CMC) (\pm expanded uncertainty)	Calibration Site
1.	Mass	Non automatic weighing instruments	Mass	In-house: ASL/QP/018/MASS Reference Methods: a) EA-4/02, b), EURAMET cg 18:2015	Standard masses Class F ₁ Traceability BS/MET/2/3/88/613 & 606 Class M ₂ Traceability BS/MET/2/3/92/732	1mg – 5 kg 5kg - 80000 kg	\pm 0.0008 g to 0.095 g \pm 0.008 kg to 2 kg	Various
		Mass standards of class M ₂ and M ₃		In-house: ASL/QP/020/MASS Reference Methods: a) EA-4/02, b) KS OIML R 111:2004	Standard masses Class F ₁ Traceability BS/MET/2/3/92/732 Class M ₁ Traceability BS/MET/2/3/92/730	1g – 200 kg	\pm 0.10 mg – 1000 mg	ASL laboratory

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	01-Oct-2020	01-Oct-2020	30-Sep-2024

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by: 
KENAS CEO/Authorized Representative

Date: 07-Jan-2022

 2022-01-22