



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION
AQUALYTIC LABORATORIES LIMITED

ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/045**

PERMANENT ADDRESS OF LABORATORY

Ramco Court, Mombasa Road
P. O. Box 4600-00506
Nairobi. Kenya.


Tel: (20) 524 96 75/ +254 721114537

Email Address: lwaweru@aqualyticlab.com

Laboratory Contact Person


Mr. Kipchumba Mugun

Approved by:



KENAS CEO/Authorized Representative

Date: 24-March-2022





KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenyaaccreditation.org Web: www.kenas.go.ke


SCHEDULE OF ACCREDITATION – TESTING

No.	Testing Field	Type of Test	Test Method	Test Object / Matrix	Test Parameter	Measurement Range	Test Location
1.	Chemistry	Potentiometric	HACH 8323 AQL/TM/C HEM/02	Water	Fluoride	0.01-3mg/L	Chemistry Laboratory
		Spectroscopy	HACH 8506 AQL/TM/C HEM/06		Copper	0.04 – 5mg/L	
			HACH 8506 AQL/TM/C HEM/04		Nickel	0.006-1 mg/L	
			HACH 8506 AQL/TM/C HEM/03		Cobalt	0.01-2mg/L	
			HACH 8506 AQL/TM/C HEM/01		Phosphorous	0.02 -2.5mg/L	
			HACH 8000 AQL/TM/C HEM/09		Chemical Oxygen Demand (COD)	0-1500mg/L	
			HACH 10129 AQL/TM/C HEM/08		Total Organic Carbon (TOC)	0-20 mg/L	
		Titrimetric	APHA standard method for chemical analysis AQL/TM/C HEM/05		Chloride	0.01-800mg/L	
			APHA standard method for chemical analysis AQL/TM/C HEM/07		Hardness	0.01 – 4000mg/L	
2	Microbiology	Enumeration	AQL/TM/BA CT/01	Water, waste water and process water	Total coliform and E.coli	≥1cfu/100ml	Microbiology Laboratory

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Assessment	30-SEPT-2019	30-SEPT-2019	29-SEPT-2023

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past.

Approved by: 
KENAS CEO/Authorized Representative

Date: 24-March 2022
