



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenas.go.ke Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

**Consolata Hospital Nkubu Laboratory
ISO 15189:2012**

Medical Laboratory Number: **KENAS/ML/69**

Date of Accreditation: **14-Dec-2018**

PERMANENT ADDRESS OF LABORATORY
NKUBU TOWN, ALONG MERU-NAIROBI HIGHWAY
P.O. BOX 205-60200
MERU, KENYA
Email: chnlaboratory@gmail.com

Laboratory Director:

Raymond Mwiraria

Technical Signatories:

1. Mary Kaburu: Haematology & Biochemistry
2. Antony Kiamba: Haematology
3. Kenneth Mwobobia: Haematology & Biochemistry
4. Peter Waithira: Haematology & Biochemistry
5. Jacob Kimathi: Haematology & Biochemistry
6. George Murithi: Haematology & Biochemistry
7. Josephine Magaju: Haematology & Biochemistry

Date of Expiry: 13-Dec-2022

Approved by:  Date: 14-Dec-2018
KENAS CEO/Authorized Representative



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
 Email: info@kenas.go.ke Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

No.	Medical Field	Examination Technique	Equipment	Specimen	Components / Analytes	Reference to Standardised Procedure	Measurement Range	Test Location
1.	Haematology	Electrical Impedance, spectrophotometry/colorimetry and floating discriminator technologies Impedance Spectrophotometry Impedance	Medonic M-SERIES, Phoenix NCC-3300	Whole Blood (EDTA)	Red Blood Cells	CLSI -H07A3 CLSI -H20-A2 CLSI -H52-A2	Medonic 0.5 -7.0X10 ¹² /L	Main Lab
					White Blood Cell	CHNL/HAEM/2 CHNL/HAEM/3	Phoenix 0.0-7.0X10 ¹² /L Medonic 0.5- 80.0X10 ⁹ /L	
					Platelet	3	Phoenix 0.0-99.9X10 ⁹ /L Medonic 30.0 - 1800.0X10 ⁹ /L	
					Haemoglobin		Phoenix 0.0-99.9X10 ⁹ /L Medonic 2.0 - 23.0 g/dL	
					Sodium	IFCC standard 2008.3- Clin Chem. 2008; 54: 559-66	30.0 - 200.0mmol/L	
					Potassium	IFCC standard 2011.2- Clin Chem Lab Med 2011; 49(9):1439-1446	20.0 - 200.0 mmol/L	
					Chloride	CLSI Approved Guideline EP5-A to assay		
						Instruction manual for NEO- ISE TYPE A Electrolyte Analyser		
						CHNL/CHEM/1 CHNL/CHEM/2		
					2.	Biochemistry	Spectrophotometry, Potentiometric, Colorimetric (Two point Rate, & Multiple point Rate),Immuno - rate	

Original Date of Accreditation: 14-Dec-2018

This schedule is issued subject to the terms and conditions of KENAS Accreditation.

Approved by: *Macharia Chesire* Date: 14-Dec-2018
 KENAS CEO/Authorized Representative