



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenas.go.ke Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

**North Kinangop Catholic Hospital Laboratory
ISO 15189:2012**

Medical Laboratory Number: **KENAS/ML/68**

Date of Accreditation: **14-Jan-2019**

PERMANENT ADDRESS OF LABORATORY

ALONG NJABINI-OLKALOU ROAD, NYANDARUA COUNTY, NORTH KINANGOP
LOCATION

P.O. BOX 88-20318

NORTH KINANGOP, KENYA

Email: northkinangoplab@gmail.com

Laboratory Director:

Fr. Borsa Sandro

Technical Signatories:

Robert Mbugua: Chemistry

Moses Njoroge: Heamatology

Date of Expiry: 13-Jan-2023

Approved by: _____

Martin Chesire
KENAS CEO/Authorized Representative

Date: 14-Jan-2019



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenas.go.ke Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

No	Medical Field	Examination Technique	Equipment	Specimen	Components / Analytes	Reference to Standardised Procedure	Measurement Range	Test Location
1.	Haematology	Electrical impedance	Coulter AC.T diff and AC.T diff 2	Whole Blood	WBC	CLSI H20-A2 NKHL-TECH-040	0.01-10.5 *10 ³ cells/ μ l	Main Lab
		Electrical impedance	Coulter AC.T diff and AC.T diff 2		RBC	CLSI H52-A2 NKHL-TECH-040	0.01-8.0 x10 ⁶ cells/ μ l	
		Photometric	Coulter AC.T diff and AC.T diff 2		HB	CLSI H15-A3 NKHL-TECH-040	1.1-25 g/dl	
		Electrical impedance	Coulter AC.T diff and AC.T diff 2		PLT	ICSH expert panel cytology NKHL-TECH-040	7-1000*10 ³ cells/ μ l	
2.	Biochemistry	Spectrophotometry and Ion Selection Electrode for electrolytes	ILAB Analyser	Serum/plasma	Urea	Ilab Aries and Ilab 300 plus application manual NKHL-TECH-054	0.2-45.8mmo/l	Main Lab
					Creatinine		5.0-2917 μ mol/l	
					Calcium		0.05-25 mmol/l	
					Sodium		20-200 mmol/l	
					Chloride		25-200 mmol/l	
					Potassium		0.2-40 mmol/l	
					ALT		4.0-1600 U/L	
					AST		6.0-1600 U/L	
					TP		0.1 -110g/l	
					D Bil		1.5-222 μ mol/l	
					T Bil		1.7-513 μ mol/l	
					HDL		0.03-6.5 mmol/l	
					TRIG		0.07-11.3 mmol/l	
					CHOL		0.1-23.3 mmo/l	
URIC	6-1488 μ mol/l							

Original Date of Accreditation: 14-Jan-2019

This schedule is issued subject to the terms and conditions of KENAS Accreditation.

Approved by:

Maechin Chesire
KENAS CEO/Authorized Representative

Date: 14-Jan-2019