



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenyaaccreditation.org Web: www.kenyaaccreditation.org

SCHEDULE OF ACCREDITATION

MOMBASA PETROGAS SERVICES

ISO 17025: 2017

Testing Laboratory Number: KENAS/ACCR/CL/12

Date of Accreditation: 30-SEP-2019

PERMANENT ADDRESS OF LABORATORY

Fimbo Footware Building, Digo Road & Shimanzi MSE, Shed No. 16, Shimanzi Road,
P. O. Box 89465-80100 GPO Digo Road Mombasa.
Tel: +254 706 514 876 / 732 424 320
Email: msapetrogas@gmail.com / info@mombasapetrogas.co.ke
Web: www.mombasapetrogas.co.ke

Technical Manager:

Austin Wise Katuu

Technical Signatories:

Pressure

1. Dennis Kalama

Volume & Flow

1. Dennis Oyugi Odhiambo

Date of expiry: 29-SEP-2023

Approved by: _____

[Signature]
KENAS CEO/Authorized Representative

Date: 30-SEP-2019

[Signature] 2019-09-30



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenyaaccreditation.org Web: www.kenyaaccreditation.org

SCHEDULE

Calibration (17025)	No.	Calibration Field	Calibration Object	Measurand	Calibration Method	Standard	(Measurement Point / Range ± expanded uncertainty)	Calibration Site
	1.	Volume	Horizontal above ground and underground tanks	Volume (0.5 – 50000 litres)	Internal method: Gauge calibration procedure MPSL/P/103 Tank calibration procedure MPSL/P/102 (Ref - MPMS)	Standard prover tanks	50000 ± 30 litres	Various
	2.	Flow	Positive displacement flow meters Micro motion meters Accuload meters	Flow (15 – 2000 litres/minute)	Internal method: Meter calibration procedure MPSL/P/101 (Ref - MPMS)	Standard prover tanks and stop watch	$Q_{min} = 15 \pm 0.03$ litres/Min $Q_r = 150 \pm 0.4$ litres/Min $Q_{max} = 3000 \pm 4$ litres/Min	
	3.	Pressure	Industrial pressure gauges Bourdon tube pressure gauges	Pressure (0.1 – 600 Bar) Pressure (1 – 10 kg/cm ²)	Internal method: Test gauge method MPSL/P/103 (Ref - DKD-R 6-1)	Test gauge S/No 211h1235022 Bourdon tube S/No 406/1	$1 - 600 \text{ bar} \pm 0.5 \%$ $1 - 10 \text{ kg/cm}^2 \pm 0.8 \%$	

Original Date of Accreditation: **30-SEP-2019**

This schedule is issued subject to the terms and conditions of KENAS Accreditation

Approved by: _____ Date: **30-SEP-2019** *[Signature]* *2mg-09-30*
 KENAS CEO/Authorized Representative