



**KENYA ACCREDITATION SERVICE**

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya  
Email: [info@kenyaaccreditation.org](mailto:info@kenyaaccreditation.org) Web: [www.kenas.go.ke](http://www.kenas.go.ke)

**SCHEDULE OF ACCREDITATION**

**VIVO ENERGY UGANDA**

**ISO/IEC 17025:2017**

Testing Laboratory Number: **KENAS/TL/062**

**Date of Accreditation: 10-March 2020**

**PERMANENT ADDRESS OF LABORATORY**

PLOT 9/11, 7<sup>th</sup> Street Industrial Area,  
P.O Box 7082  
Kampala. Uganda.

Telephone Number: +256 312 210010 / +256 772 754028

Email Address: [Allan.Kibaya@vivoenergy.com](mailto:Allan.Kibaya@vivoenergy.com)

**Website: [www.vivoenergy.com](http://www.vivoenergy.com)**

**Technical Manager:**

Allan Kibaya

**Technical Signatories:**

1. Jackson Kilimani
2. Martin Kabanda
3. Josephine Nakaya

**Date of Expiry: 09-March-2024**

Approved by: \_\_\_\_\_

*[Signature]*  
**KENAS CEO/Authorized Representative**

Date: 10-Mar-2020

*[Signature]*  
10-03-20



### KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: [info@kenyaaccreditation.org](mailto:info@kenyaaccreditation.org) Web: [www.kenas.go.ke](http://www.kenas.go.ke)

### SCHEDULE OF ACCREDITATION

No.	Testing Field	Type of Test	Test Method	Test Object / Matrix	Test Parameter	Measurement Range	Test Location
01	Chemistry	Physical - Volatility	ASTM D 86-18	Gasoline and Diesel	Distillation	400degC Max	VEU Laboratory
02		Physical - Gravimetric	ASTM D1298-12b		Density	0.775 to 0.840 kg/m <sup>3</sup> @15degC	
			ASTM D 4052-18a			0.6800 to 0.9700 kg/m <sup>3</sup>	
03		Physical - Volatility	ASTM D93A-18 Method A	Diesel	Flash Point	19 to 400degC	

**Date of Accreditation: 10-March 2020**

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past.

Approved by: \_\_\_\_\_

*[Signature]*  
KENAS CEO/Authorized Representative

Date: 10-Mar-2020

*[Signature]* 2020-03-20