



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenas.go.ke Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

KENYA AIRWAYS LIMITED

ISO/IEC 17025:2005

Testing Laboratory Number: **KENAS/CL/26**

Date of Accreditation: 13-Nov-2013

Re-Accreditation Decision: 14-Dec-2016

PERMANENT ADDRESS OF LABORATORY

Physical Address

P. O. Box 19002- 00501

Kenya Airways, Technical Department- Calibration Centre

Tel: +254 20 642 2451

Fax: +254 20 823 488

Email: Kenya-airways.com

Technical Manager:

Eng. Joel Kiprono Mibey

Technical Signatories:

1. Leah Andeso
2. Noel Kurgat

Date of expiry: 12-Nov-2019

Approved by:

Date: 14-Dec-2016

KENAS CEO/Authorized Representative



KENYA ACCREDITATION SERVICE


P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
 Email: info@kenas.go.ke Web: www.kenas.go.ke

SCHEDULE

No	Calibration Field	Calibration Object	Measurand	Calibration Method	Standard	Calibration and Measurement Capability (CMC) (Measurement Point /Range ± expanded uncertainty (k=2, level of confidence 95%))	Calibration Site
01	Force	Tensiometer Tester	Tension	KQ/TE/2/TPM Ref: ISO 7500 Traceability; KEBS Cert. No. BS/MET/3/3/45/163	Cable Tensiometer Tester	(0-5000N) ± 0.59%	Kenya Airways Calibration Centre
02	Torque	Torque Wrenches	Torque	KQ/TE/2/TPM Ref: ISO 6789 Traceability; TEAM TORQUE INC Cert. No. 04162015-508222041	Transducer	(110-1100Nm) ± 1.3Nm	
				KQ/TE/2/TPM Ref: ISO 6789 Traceability; EDUARD WILLE GmbH & Co. KG Cert. No. 0990093		(2-100Nm) ± 0.35%	
				KQ/TE/2/TPM Ref: ISO 6789 Traceability; EDUARD WILLE GmbH & Co. KG Cert. No. 0990093		(150-3000Nm) ± 0.52%	
03	Pressure	Bourdon Tube Vacuum and Pressure Gauges	Pressure	KQ/TE/2/TPM Ref: BS 837-1 Traceability; KEBS Cert. No. BS/MET/5/3/72/196	Digital Pressure Calibrator	(0-7bar) ± 0.14bar	
				KQ/TE/2/TPM Ref: BS 837-1 Traceability; KEBS Cert. No. BS/MET/5/3/72/194		(0-70bar) ± 0.02bar	
				KQ/TE/2/TPM Ref: BS 837-1 Traceability; KEBS Cert. No. BS/MET/5/3/72/195		(0-700bar) ± 0.15bar	

Original Date of Accreditation: 13-Nov 2013

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past.

Approved by:  Date: 14-Dec-2016
KENAS CEO/Authorized Representative