



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

No	Calibration Field	Calibration Object	Measurand	Calibration Method	Standard	Calibration and Measurement Capability (CMC) (Measurement Point /Range ± expanded uncertainty)	Calibration Site
01	Mass	Non Automatic Weighing Balances	Mass (1mg to 5kg)	Comparison in air Internal Method: ES/BCP/01 Reference Method: KS OIML R76 TRACEABILITY:KEBS:BS/M ET/2/3/84/640	F ₁ MASSES	1mg to 1000mg ±0.03mg 1g to 2000g ± 3.0mg 2kg to 5 Kg ± 8.0mg	Various
			Mass (5kg to 500kg)	Comparison in air Internal Method: ES/BCP/01 Reference Method: KS OIML R76 TRACEABILITY:KEBS:BS/M ET/2/3/84/72	M ₂ MASSES	5kg to 500kg ± 1g	

Original Date of Accreditation: 22-Feb-2012

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past.

Approved by:  Date: 16-Mar-2018

KENAS CEO/Authorized Representative