



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

UNGA CENTRAL LABORATORY ISO/IEC 17025:2005

Testing Laboratory Number: **KENAS/TL/25**

Date of Accreditation: 13-Nov-2013

Re-Accreditation Decision: 08-Jul-2016

PERMANENT ADDRESS OF LABORATORY

Commercial Street, Industrial Area -Nairobi

P. O. BOX 30386-00100

NAIROBI, KENYA

Tel: +254 203933000

Email: marym@unga.com

Technical Manager:

1. Mary Mureithi

Technical Signatories:

1. Judith Ngugi
2. Moses Kihui
3. Densu Kuria

Date of expiry: 12-Nov-2019

Approved by:

Date: 14-Dec-2016

KENAS CEO/Authorized Representative



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
 Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE

No.	Testing Field	Type of Test	Test Method	Test Object / Matrix	Test Parameter	Measurement Range	Test Location
01	Chemistry	Gravimetric	Internal Method: UCL/TM/05 Reference Method: ISO 5984:2002	Maize products, Wheat products, Feed stuffs and other cereals	Ash	0.01 to 100 % by wt	Nairobi Chemistry Laboratory
		Titrimetric	Internal Method: UCL/TM/06 Reference Method: ISO 6492:1999		Oil	1 to 70%by wt	
		Titrimetric	Internal Method: UCL/TM/02 Reference Method: ISO 5498:1981		Fibre	0.01 to 50% by wt	
		Gravimetric	Internal Method: UCL/TM/01 Reference Method: ISO 6496-1999		Moisture	0.1 to 30% by wt	
		Titrimetric	Internal Method: UCL/TM/07 Reference Method: ISO 5983-1:2005		Protein	0.5 to 90% by wt	
		Titrimetric	Internal Method: UCL/TM/04 Reference Method: KS ISO6490-1985		Calcium	0.05 to 80% by wt	
		Gravimetric	Internal Method: UCL/TM/03				

Original Date of Accreditation: 13th November 2013

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past.

Approved by: _____ Date: 14-Dec-2016

KENAS CEO/Authorized Representative